

## Certificate of Compliance

**Product Description** : 14 ml Round-Bottom Tubes, 17x95mm Style with Snap Cap, Sterile

**Catalog Number** : 40002004 / 40002008

**Lot Number** : 14400920

**Manufacture Date** : 04/2023

**Expiry Date** : 04/2026

**Visual Attribute** : PASS

**Functional Test** : PASS

**Sterilization** : Sterilization was performed by gamma irradiation according to ISO 11137-1, the Sterilization does <SAL 10<sup>-6</sup>> establishing by ISO 1137-2

**Cytotoxicity Test** : Cytotoxicity test is conducted to qualify all material Resins using USP <87> and or ISO 10993-5 standards for Cytotoxicity and have been shown to be nontoxic.

**RNase/DNase/DNA Test** : This product has been tested and is free of any detectable RNase/DNase/DNA contamination

**MEA Test** : 1 Cell to 2 Cell within 24 Hours  
1 Cell to Expanded blastocyst within 96 hours.  
100% achieved.(Embryotech USA-AT-6579-0821)

ARCK TECHNOLOGIES INNOVATIONS PVT LTD. hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented.

**Date:** 20<sup>th</sup> May 2023

Authorized Signatory

