

Certificate of Compliance

Product Description : 15ml Conical Centrifuge Tubes (Individually Packed Sterile)

Catalog Number : 40002005

Lot Number : 15400706

Manufacture Date : 04/2023

Expiry Date : 04/2026

Visual Attribute : PASS

Functional Test : PASS

Sterilization : Sterilization was performed by gamma irradiation according to ISO11137-1, the Sterilization does <SAL 10⁻⁶> establishing by ISO 1137-2

Cytotoxicity Test : Cytotoxicity test is conducted to qualify all material Resins using USP <87> and or ISO 10993-5 standards for Cytotoxicity and have been shown to be nontoxic

RNase/DNase/DNA : This product has been tested and is free of any detectable Test RNase /DNase/DNA contamination

ARCK TECHNOLOGIES INNOVATIONS PVT LTD. hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented.

Date : 20th May 2023



Authorized Signature

