



ARCK
TECHNOLOGIES[®]
INNOVATIONS PVT LTD

Certificate of Compliance

Product Description : Culture Dish – Pack of 5, Sterile

Catalog Number : 40002014

Lot Number : CD400123

Manufacture Date : 04/2023

Expiry Date : 04/2026

Visual Attribute : PASS

Functional Test : PASS

Sterilization : Sterilization was performed by gamma irradiation according to ISO 11137-1, the Sterilization does <SAL 10-6> establishing by ISO 1137-2

Cytotoxicity Test : Cytotoxicity test is conducted to qualify all material Resins using USP <87> and or ISO 10993-5 standards for Cytotoxicity and have been shown to be nontoxic.

RNase/DNase/DNA Test : This product has been tested and is free of any detectable RNase/DNase/DNA contamination

MEA Test : 1 Cell to 2 Cell within 24 Hours
1 Cell to Expanded blastocyst within 96 hours.
100% achieved.(Embryotech USA-AT-7322-1221)

Pyrogen/ Endotoxin Test : Non-pyrogenic tested according to the principles of the LAL test.
Test Result: <0.03125EU/Device. (ELI Accession No: E1842-1221AT)

ARCK TECHNOLOGIES INNOVATIONS PVT LTD. hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented.

Date: 20th May 2023

Authorized Signatory

