

Certificate of Compliance

Product Description : 1000µl UNIVERSAL GRAD MICRO TIPS, Sterile

Catalog Number : **40002041**

Lot Number : 00082

Manufacture Date : 06/2022

Expiry Date : 06/2027

Visual Attribute : PASS

Functional Test : PASS

Sterilization : Sterilization was performed by gamma irradiation according to ISO 11137-1, the Sterilization does <SAL 10-6> establishing by ISO 1137-2

Cytotoxicity Test : Cytotoxicity test is conducted to qualify all material Resins using USP <87> and or ISO 10993-5 standards for Cytotoxicity and have been shown to be nontoxic

RNase/DNase/DNA Test : This product has been tested and is free of any detectable RNase/DNase/DNA contamination

ARCK TECHNOLOGIES INNOVATIONS PVT LTD hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented.

Date : 26th July 2022


Authorized Signature

